

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN052319\
 Data File : VN055759.D
 Acq On : 23 May 2019 17:36
 Operator : JC/SP
 Sample : K3008-18
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 COMP-20

Quant Time: May 24 07:21:27 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N052319W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 24 06:41:54 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	461375	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	631848	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	566161	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	214051	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	209101	43.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.78%	
35) Dibromofluoromethane	7.59	113	191792	48.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.02%	
50) Toluene-d8	10.09	98	745764	48.67	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.34%	
62) 4-Bromofluorobenzene	12.40	95	242140	46.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.16%	
Target Compounds						
16) Acetone	3.82	43	11984	5.722	ug/l	99
20) Methylene Chloride	4.54	84	27215	6.000	ug/l	96
43) Isopropyl Acetate	8.17	43	14101	1.577	ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN052319\
Data File : VN055759.D
Acq On : 23 May 2019 17:36
Operator : JC/SP
Sample : K3008-18
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampled :
COMP-20

Quant Time: May 24 07:21:27 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N052319W.M
Quant Title : SW846 8260
QLast Update : Fri May 24 06:41:54 2019
Response via : Initial Calibration

