

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN052319\
 Data File : VN055760.D
 Acq On : 23 May 2019 18:01
 Operator : JC/SP
 Sample : K3008-07
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 COMP-19

Quant Time: May 24 07:23:17 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N052319W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 24 06:41:54 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	438786	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	606585	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	532448	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	192011	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	200248	44.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.38%	
35) Dibromofluoromethane	7.59	113	185295	48.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.64%	
50) Toluene-d8	10.09	98	715480	48.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.28%	
62) 4-Bromofluorobenzene	12.40	95	226657	44.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.86%	
Target Compounds						
16) Acetone	3.82	43	10097	5.069	ug/l	97
20) Methylene Chloride	4.55	84	79241	18.369	ug/l	97
43) Isopropyl Acetate	8.17	43	14044	1.636	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN052319\
Data File : VN055760.D
Acq On : 23 May 2019 18:01
Operator : JC/SP
Sample : K3008-07
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
COMP-19

Quant Time: May 24 07:23:17 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N052319W.M
Quant Title : SW846 8260
QLast Update : Fri May 24 06:41:54 2019
Response via : Initial Calibration

