


Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN052319\
Data File : VN055752.D
Acq On : 23 May 2019 14:28
Operator : JC/SP
Sample : BFB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N052319W.M
Title : SW846 8260
Last Update : Fri May 24 06:41:54 2019



AutoFind: Scans 3304, 3305, 3306; Background Corrected with Scan 3291

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.8	6554	PASS
75	95	30	60	46.6	18168	PASS
95	95	100	100	100.0	39013	PASS
96	95	5	9	6.5	2544	PASS
173	174	0.00	2	0.3	105	PASS
174	95	50	100	98.7	38522	PASS
175	174	5	9	7.3	2813	PASS
176	174	95	101	96.3	37104	PASS
177	176	5	9	6.5	2413	PASS