

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052323\
 Data File : VN077806.D
 Acq On : 23 May 2023 10:57
 Operator : JC\MD
 Sample : VN0523MBL01
 Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0523MBL01

Quant Time: May 24 02:52:52 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 04:07:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	383746	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	681800	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.876	117	586401	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	194194	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	248229	43.232	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	86.460%
35) Dibromofluoromethane	8.171	113	223694	50.314	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.620%
50) Toluene-d8	10.576	98	832474	48.646	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.300%
62) 4-Bromofluorobenzene	12.859	95	247558	40.960	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	81.920%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052323\
Data File : VN077806.D
Acq On : 23 May 2023 10:57
Operator : JC\MD
Sample : VN0523MBL01
Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0523MBL01

Quant Time: May 24 02:52:52 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
Quant Title : SW846 8260
QLast Update : Tue May 16 04:07:42 2023
Response via : Initial Calibration

