

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052323\
 Data File : VN077810.D
 Acq On : 23 May 2023 12:43
 Operator : JC\MD
 Sample : 02880-02DL 5X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RE132D5-20230517DL

Quant Time: May 24 02:53:51 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 04:07:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	383010	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	682591	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	572011	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	184386	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	249268	43.496	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	87.000%
35) Dibromofluoromethane	8.171	113	217372	48.836	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.680%
50) Toluene-d8	10.571	98	819898	47.856	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	95.720%
62) 4-Bromofluorobenzene	12.859	95	256249	42.349	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	84.700%
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	2.136	85	1213	0.278	ug/l	85
9) 1,1,2-Trichlorotrifluo...	4.377	101	8852	2.088	ug/l #	47
12) 1,1-Dichloroethene	4.347	96	1924	0.478	ug/l #	79
27) cis-1,2-Dichloroethene	7.500	96	2101	0.387	ug/l	66
30) Chloroform	7.965	83	2481	0.265	ug/l #	81
44) Trichloroethene	9.359	130	159085	32.702	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052323\
Data File : VN077810.D
Acq On : 23 May 2023 12:43
Operator : JC\MD
Sample : 02880-02DL 5X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RE132D5-20230517DL

Quant Time: May 24 02:53:51 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
Quant Title : SW846 8260
QLast Update : Tue May 16 04:07:42 2023
Response via : Initial Calibration

