

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN052323\  
 Data File : VN077813.D  
 Acq On : 23 May 2023 13:55  
 Operator : JC\MD  
 Sample : 02880-07  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 10 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 RW8-MW01D1-20230518

Quant Time: May 24 02:54:21 2023  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N051523W.M  
 Quant Title : SW846 8260  
 QLast Update : Tue May 16 04:07:42 2023  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon  | Response | Conc     | Units | Dev(Min) |
|-----------------------------|--------|-------|----------|----------|-------|----------|
| -----                       |        |       |          |          |       |          |
| Internal Standards          |        |       |          |          |       |          |
| 1) Pentafluorobenzene       | 8.230  | 168   | 375809   | 50.000   | ug/l  | 0.00     |
| 34) 1,4-Difluorobenzene     | 9.106  | 114   | 671157   | 50.000   | ug/l  | 0.00     |
| 63) Chlorobenzene-d5        | 11.871 | 117   | 576684   | 50.000   | ug/l  | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 13.800 | 152   | 183213   | 50.000   | ug/l  | 0.00     |
|                             |        |       |          |          |       |          |
| System Monitoring Compounds |        |       |          |          |       |          |
| 33) 1,2-Dichloroethane-d4   | 8.583  | 65    | 251118   | 44.659   | ug/l  | 0.00     |
| Spiked Amount               | 50.000 | Range | 74 - 125 | Recovery | =     | 89.320%  |
| 35) Dibromofluoromethane    | 8.171  | 113   | 211426   | 48.309   | ug/l  | 0.00     |
| Spiked Amount               | 50.000 | Range | 75 - 124 | Recovery | =     | 96.620%  |
| 50) Toluene-d8              | 10.571 | 98    | 804667   | 47.767   | ug/l  | 0.00     |
| Spiked Amount               | 50.000 | Range | 86 - 113 | Recovery | =     | 95.540%  |
| 62) 4-Bromofluorobenzene    | 12.853 | 95    | 260311   | 43.753   | ug/l  | 0.00     |
| Spiked Amount               | 50.000 | Range | 83 - 123 | Recovery | =     | 87.500%  |
|                             |        |       |          |          |       |          |
| Target Compounds            |        |       |          |          |       | Qvalue   |
| 44) Trichloroethene         | 9.359  | 130   | 2248     | 0.470    | ug/l  | 40       |
| -----                       |        |       |          |          |       |          |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN052323\  
Data File : VN077813.D  
Acq On : 23 May 2023 13:55  
Operator : JC\MD  
Sample : 02880-07  
Misc : 5.0mL/MSVOA\_N/WATER  
ALS Vial : 10 Sample Multiplier: 1

Instrument :  
MSVOA\_N  
ClientSampleId :  
RW8-MW01D1-20230518

Quant Time: May 24 02:54:21 2023  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N051523W.M  
Quant Title : SW846 8260  
QLast Update : Tue May 16 04:07:42 2023  
Response via : Initial Calibration

