

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052323\
 Data File : VN077816.D
 Acq On : 23 May 2023 15:06
 Operator : JC\MD
 Sample : 02880-17
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TT161S1-20230519

Manual Integrations
 APPROVED

Reviewed By : John Carlone 05/25/2023
 Supervised By : Mahesh Dadoda 05/25/2023

Quant Time: May 24 02:54:47 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 04:07:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	376464	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	677507	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	601322	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	194844	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	252802	44.880	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	89.760%		
35) Dibromofluoromethane	8.177	113	218915	49.551	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	99.100%		
50) Toluene-d8	10.576	98	840019	49.398	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	98.800%		
62) 4-Bromofluorobenzene	12.853	95	264081	43.971	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	87.940%		
Target Compounds					Qvalue	
12) 1,1-Dichloroethene	4.359	96	2775	0.701	ug/l #	86
16) Acetone	4.436	43	4389m	2.477	ug/l	
24) 1,1-Dichloroethane	6.577	63	22336	2.674	ug/l	94
30) Chloroform	7.977	83	2991	0.326	ug/l #	81
44) Trichloroethene	9.365	130	4268	0.884	ug/l	56
64) Tetrachloroethene	11.112	164	1940	0.461	ug/l #	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052323\
Data File : VN077816.D
Acq On : 23 May 2023 15:06
Operator : JC\MD
Sample : 02880-17
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TT161S1-20230519

Quant Time: May 24 02:54:47 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
Quant Title : SW846 8260
QLast Update : Tue May 16 04:07:42 2023
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : John Carlone 05/25/2023
Supervised By : Mahesh Dadoda 05/25/2023

