

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052323\
 Data File : VN077833.D
 Acq On : 24 May 2023 02:14
 Operator : JC\MD
 Sample : PB152982TB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB152982TB

Quant Time: May 24 05:58:38 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 04:07:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	407435	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.112	114	726315	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	662854	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	218337	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	267465	43.874	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	87.740%		
35) Dibromofluoromethane	8.177	113	231291	48.835	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	97.660%		
50) Toluene-d8	10.576	98	912952	50.080	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	100.160%		
62) 4-Bromofluorobenzene	12.859	95	288662	44.834	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	89.660%		
Target Compounds						
					Qvalue	
16) Acetone	4.441	43	28475	14.846	ug/l #	86
20) Methylene Chloride	5.294	84	24332	3.018	ug/l #	94
37) Ethyl Acetate	7.571	43	66332	10.743	ug/l	96
43) Isopropyl Acetate	8.694	43	280731	27.835	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052323\
Data File : VN077833.D
Acq On : 24 May 2023 02:14
Operator : JC\MD
Sample : PB152982TB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB152982TB

Quant Time: May 24 05:58:38 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
Quant Title : SW846 8260
QLast Update : Tue May 16 04:07:42 2023
Response via : Initial Calibration

