

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052323\
 Data File : VN077838.D
 Acq On : 24 May 2023 04:14
 Operator : JC\MD
 Sample : 02892-05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-31

Quant Time: May 24 05:59:43 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 04:07:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.236	168	404979	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	720403	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	673501	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	236474	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	261851	43.213	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	86.420%
35) Dibromofluoromethane	8.177	113	227177	48.360	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.720%
50) Toluene-d8	10.571	98	905071	50.055	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.100%
62) 4-Bromofluorobenzene	12.853	95	300294	47.023	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.040%
Target Compounds						
					Qvalue	
16) Acetone	4.448	43	48737	25.565	ug/l #	86
20) Methylene Chloride	5.289	84	24242	3.029	ug/l	90
37) Ethyl Acetate	7.571	43	64365	10.510	ug/l	98
43) Isopropyl Acetate	8.700	43	336325	33.621	ug/l #	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052323\
Data File : VN077838.D
Acq On : 24 May 2023 04:14
Operator : JC\MD
Sample : 02892-05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-31

Quant Time: May 24 05:59:43 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
Quant Title : SW846 8260
QLast Update : Tue May 16 04:07:42 2023
Response via : Initial Calibration

