

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052323\
 Data File : VN077839.D
 Acq On : 24 May 2023 04:37
 Operator : JC\MD
 Sample : 02892-08
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-33

Quant Time: May 24 05:59:54 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 04:07:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.236	168	388191	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	719821	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.876	117	653514	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	233790	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	261290	44.985	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	89.980%
35) Dibromofluoromethane	8.177	113	223497	47.615	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.220%
50) Toluene-d8	10.577	98	902442	49.950	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.900%
62) 4-Bromofluorobenzene	12.859	95	293509	45.998	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.000%
Target Compounds						
					Qvalue	
16) Acetone	4.448	43	47504	25.996	ug/l	97
20) Methylene Chloride	5.283	84	25004	3.416	ug/l	84
37) Ethyl Acetate	7.565	43	65507	10.705	ug/l #	95
43) Isopropyl Acetate	8.694	43	321310	32.146	ug/l #	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052323\
Data File : VN077839.D
Acq On : 24 May 2023 04:37
Operator : JC\MD
Sample : 02892-08
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-33

Quant Time: May 24 05:59:54 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
Quant Title : SW846 8260
QLast Update : Tue May 16 04:07:42 2023
Response via : Initial Calibration

