

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052323\
 Data File : VN077841.D
 Acq On : 24 May 2023 05:25
 Operator : JC\MD
 Sample : 02892-14
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-29

Quant Time: May 24 06:00:16 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 04:07:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.236	168	387395	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	708363	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	642380	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	210337	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	260992	45.026	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	90.060%
35) Dibromofluoromethane	8.177	113	223398	48.363	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.720%
50) Toluene-d8	10.571	98	886062	49.836	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.680%
62) 4-Bromofluorobenzene	12.859	95	285201	45.419	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.840%
Target Compounds						
					Qvalue	
16) Acetone	4.442	43	48925	26.828	ug/l	93
20) Methylene Chloride	5.295	84	22374	2.851	ug/l #	82
37) Ethyl Acetate	7.571	43	60344	10.021	ug/l	98
43) Isopropyl Acetate	8.694	43	308030	31.316	ug/l #	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052323\
Data File : VN077841.D
Acq On : 24 May 2023 05:25
Operator : JC\MD
Sample : 02892-14
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-29

Quant Time: May 24 06:00:16 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
Quant Title : SW846 8260
QLast Update : Tue May 16 04:07:42 2023
Response via : Initial Calibration

