

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052323\
 Data File : VN077842.D
 Acq On : 24 May 2023 05:49
 Operator : JC\MD
 Sample : 02894-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-40A

Quant Time: May 24 06:34:59 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 04:07:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	385282	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.112	114	697687	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	641277	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	215127	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	257753	44.711	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	89.420%
35) Dibromofluoromethane	8.171	113	219120	48.163	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.320%
50) Toluene-d8	10.571	98	874634	49.946	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.900%
62) 4-Bromofluorobenzene	12.859	95	279497	45.192	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.380%
Target Compounds						
					Qvalue	
16) Acetone	4.447	43	37344	20.590	ug/l	98
20) Methylene Chloride	5.294	84	21812	2.754	ug/l #	90
37) Ethyl Acetate	7.571	43	63366	10.684	ug/l	97
43) Isopropyl Acetate	8.694	43	291846	30.125	ug/l #	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052323\
Data File : VN077842.D
Acq On : 24 May 2023 05:49
Operator : JC\MD
Sample : 02894-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-40A

Quant Time: May 24 06:34:59 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
Quant Title : SW846 8260
QLast Update : Tue May 16 04:07:42 2023
Response via : Initial Calibration

