

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052323\
 Data File : VN077843.D
 Acq On : 24 May 2023 06:13
 Operator : JC\MD
 Sample : 02894-05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-50A

Quant Time: May 24 06:35:11 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 04:07:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	389494	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	705227	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	645352	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	221017	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	257224	44.137	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	88.280%
35) Dibromofluoromethane	8.177	113	225735	49.087	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.180%
50) Toluene-d8	10.571	98	881069	49.776	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.560%
62) 4-Bromofluorobenzene	12.853	95	285095	45.604	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.200%
Target Compounds						
					Qvalue	
16) Acetone	4.442	43	40538	22.109	ug/l	93
20) Methylene Chloride	5.277	84	21519	2.638	ug/l #	93
37) Ethyl Acetate	7.577	43	63372	10.570	ug/l	97
43) Isopropyl Acetate	8.694	43	304963	31.142	ug/l #	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052323\
Data File : VN077843.D
Acq On : 24 May 2023 06:13
Operator : JC\MD
Sample : 02894-05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-50A

Quant Time: May 24 06:35:11 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
Quant Title : SW846 8260
QLast Update : Tue May 16 04:07:42 2023
Response via : Initial Calibration

