

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052324\
 Data File : VN082311.D
 Acq On : 23 May 2024 10:54
 Operator : JC\MD
 Sample : VN0523WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0523WBL01

Quant Time: May 24 04:13:55 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 16 05:21:06 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	171251	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	306110	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	288939	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	122707	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	144260	53.349	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	106.700%
35) Dibromofluoromethane	8.171	113	101010	53.642	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	107.280%
50) Toluene-d8	10.571	98	395178	54.738	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	109.480%
62) 4-Bromofluorobenzene	12.847	95	131378	50.985	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	101.980%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052324\
Data File : VN082311.D
Acq On : 23 May 2024 10:54
Operator : JC\MD
Sample : VN0523WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0523WBL01

Quant Time: May 24 04:13:55 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
Quant Title : SW846 8260
QLast Update : Thu May 16 05:21:06 2024
Response via : Initial Calibration

