

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052324\
 Data File : VN082314.D
 Acq On : 23 May 2024 12:16
 Operator : JC\MD
 Sample : P2554-01 50X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 AUD-1041

Quant Time: May 24 04:16:08 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 16 05:21:06 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	152937	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	274376	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	251900	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	104305	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	127923	52.972	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 105.940%			
35) Dibromofluoromethane	8.165	113	89650	53.115	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 106.240%			
50) Toluene-d8	10.571	98	347530	53.706	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 107.420%			
62) 4-Bromofluorobenzene	12.853	95	115470	49.994	ug/l	0.00
Spiked Amount 50.000	Range 64 - 133		Recovery = 99.980%			
Target Compounds						
16) Acetone	4.436	43	6424	6.485	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052324\
Data File : VN082314.D
Acq On : 23 May 2024 12:16
Operator : JC\MD
Sample : P2554-01 50X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
AUD-1041

Quant Time: May 24 04:16:08 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
Quant Title : SW846 8260
QLast Update : Thu May 16 05:21:06 2024
Response via : Initial Calibration

