

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052324\
 Data File : VN082321.D
 Acq On : 23 May 2024 15:06
 Operator : JC\MD
 Sample : P2556-01
 Misc : 5.01g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 COMP-1

Quant Time: May 24 04:19:11 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 16 05:21:06 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.223	168	170157	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	301004	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	282385	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	154232	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.576	65	135934	50.593	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 101.180%			
35) Dibromofluoromethane	8.165	113	94994	51.303	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 102.600%			
50) Toluene-d8	10.570	98	381351	53.719	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 107.440%			
62) 4-Bromofluorobenzene	12.853	95	151148	59.652	ug/l	0.00
Spiked Amount 50.000	Range 64 - 133		Recovery = 119.300%			
Target Compounds						Qvalue
20) Methylene Chloride	5.247	84	3661	1.646	ug/l #	73
31) Cyclohexane	8.212	56	16909	3.940	ug/l #	66
52) Toluene	10.629	92	111981	20.150	ug/l	96
67) Ethyl Benzene	11.970	91	25150	2.254	ug/l	94
68) m/p-Xylenes	12.070	106	28590	6.990	ug/l	91
69) o-Xylene	12.400	106	15556	3.926	ug/l	79
80) 1,3,5-Trimethylbenzene	13.176	105	28464	2.873	ug/l	96
84) 1,2,4-Trimethylbenzene	13.482	105	98151	9.683	ug/l	98
86) p-Isopropyltoluene	13.735	119	11180	1.145	ug/l	92
89) n-Butylbenzene	14.058	91	11144	1.270	ug/l #	64

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052324\
Data File : VN082321.D
Acq On : 23 May 2024 15:06
Operator : JC\MD
Sample : P2556-01
Misc : 5.01g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
COMP-1

Quant Time: May 24 04:19:11 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
Quant Title : SW846 8260
QLast Update : Thu May 16 05:21:06 2024
Response via : Initial Calibration

