

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052324\
 Data File : VN082326.D
 Acq On : 23 May 2024 17:07
 Operator : JC\MD
 Sample : P2562-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 FIELD-BLANK

Quant Time: May 24 04:21:27 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 16 05:21:06 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	202620	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	354992	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	337493	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	155805	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	158593	49.569	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	99.140%
35) Dibromofluoromethane	8.171	113	112432	51.486	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.980%
50) Toluene-d8	10.571	98	444249	53.062	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	106.120%
62) 4-Bromofluorobenzene	12.853	95	168999	56.554	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	113.100%
Target Compounds						
					Qvalue	
16) Acetone	4.442	43	10818	8.243	ug/l #	85
25) 2-Butanone	7.494	43	4784	2.404	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052324\
Data File : VN082326.D
Acq On : 23 May 2024 17:07
Operator : JC\MD
Sample : P2562-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
FIELD-BLANK

Quant Time: May 24 04:21:27 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
Quant Title : SW846 8260
QLast Update : Thu May 16 05:21:06 2024
Response via : Initial Calibration

