

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052324\
 Data File : VN082327.D
 Acq On : 23 May 2024 17:32
 Operator : JC\MD
 Sample : P2562-03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SY-6

Quant Time: May 24 04:21:49 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 16 05:21:06 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	202212	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	356470	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	336109	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	148430	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	160922	50.399	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	100.800%	
35) Dibromofluoromethane	8.171	113	112159	51.148	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	102.300%	
50) Toluene-d8	10.571	98	448436	53.340	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	106.680%	
62) 4-Bromofluorobenzene	12.853	95	164376	54.779	ug/l	0.00
Spiked Amount	50.000	Range 64 - 133	Recovery	=	109.560%	
Target Compounds						
16) Acetone	4.442	43	3914	2.988	ug/l	Qvalue 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052324\
Data File : VN082327.D
Acq On : 23 May 2024 17:32
Operator : JC\MD
Sample : P2562-03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SY-6

Quant Time: May 24 04:21:49 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
Quant Title : SW846 8260
QLast Update : Thu May 16 05:21:06 2024
Response via : Initial Calibration

