

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052324\
 Data File : VN082328.D
 Acq On : 23 May 2024 17:56
 Operator : JC\MD
 Sample : P2561-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SY-2R

Quant Time: May 24 04:22:11 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 16 05:21:06 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	201114	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	357177	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	341072	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	157465	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	162366	51.129	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 102.260%	
35) Dibromofluoromethane	8.171	113	113526	51.669	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 103.340%	
50) Toluene-d8	10.571	98	455927	54.123	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 108.240%	
62) 4-Bromofluorobenzene	12.853	95	172215	57.278	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	= 114.560%	
Target Compounds						
					Qvalue	
16) Acetone	4.442	43	4618	3.545	ug/l #	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052324\
Data File : VN082328.D
Acq On : 23 May 2024 17:56
Operator : JC\MD
Sample : P2561-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SY-2R

Quant Time: May 24 04:22:11 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
Quant Title : SW846 8260
QLast Update : Thu May 16 05:21:06 2024
Response via : Initial Calibration

