

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052325\
 Data File : VN086767.D
 Acq On : 23 May 2025 10:28
 Operator : JC\MD
 Sample : VN0523MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0523MBL01

Quant Time: May 24 02:29:29 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051625W.M
 Quant Title : SW846 8260
 QLast Update : Sat May 17 00:46:13 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	238191	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	441301	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	422910	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	177999	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	144540	43.915	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	87.820%
35) Dibromofluoromethane	8.165	113	132802	50.531	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.060%
50) Toluene-d8	10.565	98	562468	51.645	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	103.280%
62) 4-Bromofluorobenzene	12.847	95	198524	51.624	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	103.240%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052325\
Data File : VN086767.D
Acq On : 23 May 2025 10:28
Operator : JC\MD
Sample : VN0523MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0523MBL01

Quant Time: May 24 02:29:29 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051625W.M
Quant Title : SW846 8260
QLast Update : Sat May 17 00:46:13 2025
Response via : Initial Calibration

