

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052325\
 Data File : VN086771.D
 Acq On : 23 May 2025 12:16
 Operator : JC\MD
 Sample : Q2115-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 OUTFALL-1

Quant Time: May 24 02:31:58 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051625W.M
 Quant Title : SW846 8260
 QLast Update : Sat May 17 00:46:13 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	247918	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	461976	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	449647	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	204624	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	153536	44.818	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	89.640%
35) Dibromofluoromethane	8.165	113	139491	50.701	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.400%
50) Toluene-d8	10.565	98	582571	51.097	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	102.200%
62) 4-Bromofluorobenzene	12.847	95	216994	53.901	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	107.800%
Target Compounds						
					Qvalue	
16) Acetone	4.424	43	44881	35.023	ug/l	90
25) 2-Butanone	7.483	43	5663	2.860	ug/l	99
95) Naphthalene	15.635	128	10875	1.133	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052325\
Data File : VN086771.D
Acq On : 23 May 2025 12:16
Operator : JC\MD
Sample : Q2115-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
OUTFALL-1

Quant Time: May 24 02:31:58 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051625W.M
Quant Title : SW846 8260
QLast Update : Sat May 17 00:46:13 2025
Response via : Initial Calibration

