

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052325\
 Data File : VN086772.D
 Acq On : 23 May 2025 12:40
 Operator : JC\MD
 Sample : Q2116-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 OUTFALL-2

Quant Time: May 24 02:32:21 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051625W.M
 Quant Title : SW846 8260
 QLast Update : Sat May 17 00:46:13 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	246001	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	452072	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	437261	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	192791	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	149096	43.861	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	87.720%
35) Dibromofluoromethane	8.165	113	136599	50.737	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.480%
50) Toluene-d8	10.565	98	577501	51.762	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	103.520%
62) 4-Bromofluorobenzene	12.847	95	206691	52.467	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	104.940%
Target Compounds						
					Qvalue	
3) Chloromethane	2.365	50	1289	0.321	ug/l	97
16) Acetone	4.442	43	8986	7.067	ug/l	90
95) Naphthalene	15.635	128	4789	0.530	ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052325\
Data File : VN086772.D
Acq On : 23 May 2025 12:40
Operator : JC\MD
Sample : Q2116-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
OUTFALL-2

Quant Time: May 24 02:32:21 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051625W.M
Quant Title : SW846 8260
QLast Update : Sat May 17 00:46:13 2025
Response via : Initial Calibration

