

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN052419\
 Data File : VN055783.D
 Acq On : 24 May 2019 9:52
 Operator : JC/SP
 Sample : VN0524WBL01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0524WBL01

Quant Time: May 24 13:03:33 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA N\METHODS\624N052219W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Thu May 23 05:54:54 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	7.19	128	60767	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	8.58	114	321898	30.00	ug/l	0.00
57) Chlorobenzene-d5	11.41	117	255104	30.00	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.03	65	113557	28.59	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	95.30%
60) 4-Bromofluorobenzene	12.40	95	93880	23.66	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	78.87%
63) Toluene-d8	10.09	98	370452	31.78	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	105.93%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN052419\
Data File : VN055783.D
Acq On : 24 May 2019 9:52
Operator : JC/SP
Sample : VN0524WBL01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0524WBL01

Quant Time: May 24 13:03:33 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA N\METHODS\624N052219W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Thu May 23 05:54:54 2019
Response via : Initial Calibration

