

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN052419\
 Data File : VN055791.D
 Acq On : 24 May 2019 13:12
 Operator : JC/SP
 Sample : K2977-04 100X
 Misc : 5.61g/5mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 EP1-W-2R(17-20)

Quant Time: May 24 14:16:48 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N052319W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 24 06:41:54 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	421562	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	595306	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	549729	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	232415	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	194214	44.61	ug/l	0.00
Spiked Amount	50.000		Recovery	=	89.22%	
35) Dibromofluoromethane	7.59	113	177889	47.26	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.52%	
50) Toluene-d8	10.09	98	708832	49.10	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.20%	
62) 4-Bromofluorobenzene	12.40	95	245795	49.65	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.30%	
Target Compounds						
					Qvalue	
68) m/p-Xylenes	11.62	106	25438	3.410	ug/l	100
73) Isopropylbenzene	12.25	105	60948	1.077	ug/l	100
78) n-propylbenzene	12.59	91	188708	9.342	ug/l	99
84) 1,2,4-Trimethylbenzene	13.04	105	1119906	82.112	ug/l	99
85) sec-Butylbenzene	13.17	105	115551	5.273	ug/l #	68
86) p-Isopropyltoluene	13.29	119	121155	7.506	ug/l	99
89) n-Butylbenzene	13.61	91	76939	6.221	ug/l #	72

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN052419\
Data File : VN055791.D
Acq On : 24 May 2019 13:12
Operator : JC/SP
Sample : K2977-04 100X
Misc : 5.61µ/5mL/100µL/5.00mL/MSVOA_N/MEOH
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
EP1-W-2R(17-20)

Quant Time: May 24 14:16:48 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N052319W.M
Quant Title : SW846 8260
QLast Update : Fri May 24 06:41:54 2019
Response via : Initial Calibration

