

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN052419\
 Data File : VN055818.D
 Acq On : 25 May 2019 1:19
 Operator : JC/SP
 Sample : K3022-02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-3

Quant Time: May 25 05:39:53 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N052319W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 24 06:41:54 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	459063	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	642540	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	576420	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	198490	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	211115	44.53	ug/l	0.00
Spiked Amount	50.000		Recovery	=	89.06%	
35) Dibromofluoromethane	7.59	113	191437	47.12	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.24%	
50) Toluene-d8	10.09	98	748985	48.06	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.12%	
62) 4-Bromofluorobenzene	12.40	95	240751	45.06	ug/l	0.00
Spiked Amount	50.000		Recovery	=	90.12%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN052419\
Data File : VN055818.D
Acq On : 25 May 2019 1:19
Operator : JC/SP
Sample : K3022-02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-3

Quant Time: May 25 05:39:53 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N052319W.M
Quant Title : SW846 8260
QLast Update : Fri May 24 06:41:54 2019
Response via : Initial Calibration

