

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN052419\
 Data File : VN055835.D
 Acq On : 25 May 2019 8:24
 Operator : JC/SP
 Sample : K3026-04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 58 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-5-S

Quant Time: May 27 00:00:07 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N052319W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 24 06:41:54 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	394235	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	562716	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	511703	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	184399	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	185775	45.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.26%	
35) Dibromofluoromethane	7.59	113	170670	47.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.94%	
50) Toluene-d8	10.09	98	669247	49.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.08%	
62) 4-Bromofluorobenzene	12.40	95	218376	46.67	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.34%	
Target Compounds						
16) Acetone	3.82	43	18724	10.462	ug/l	98
20) Methylene Chloride	4.54	84	24657	6.362	ug/l	91
43) Isopropyl Acetate	8.17	43	10796	1.356	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN052419\
Data File : VN055835.D
Acq On : 25 May 2019 8:24
Operator : JC/SP
Sample : K3026-04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 58 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-5-S

Quant Time: May 27 00:00:07 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N052319W.M
Quant Title : SW846 8260
QLast Update : Fri May 24 06:41:54 2019
Response via : Initial Calibration

