

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052422\
 Data File : VN072588.D
 Acq On : 24 May 2022 15:04
 Operator : JC\MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: May 25 09:18:43 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N052422W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 25 09:15:20 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	219681	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	310219	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	232278	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.669	152	99186	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	109691	43.896	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	87.800%
35) Dibromofluoromethane	8.016	113	97196	48.341	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	96.680%
50) Toluene-d8	10.439	98	323413	42.126	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	84.260%
62) 4-Bromofluorobenzene	12.727	95	90767	36.351	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	72.700%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052422\
Data File : VN072588.D
Acq On : 24 May 2022 15:04
Operator : JC\MD
Sample : IBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: May 25 09:18:43 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N052422W.M
Quant Title : SW846 8260
QLast Update : Wed May 25 09:15:20 2022
Response via : Initial Calibration

