

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052422\
 Data File : VN072595.D
 Acq On : 24 May 2022 18:54
 Operator : JC\MD
 Sample : N2981-21
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TGMP-1

Quant Time: May 25 09:21:04 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N052422W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 25 09:15:20 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	150492	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	281202	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	320870	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	99987	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	89787	52.450	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 104.900%	
35) Dibromofluoromethane	8.022	113	78977	43.333	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 86.660%	
50) Toluene-d8	10.439	98	356938	51.291	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 102.580%	
62) 4-Bromofluorobenzene	12.727	95	135289	59.773	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 119.540%	
Target Compounds						
16) Acetone	4.287	43	2965	3.524	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052422\
Data File : VN072595.D
Acq On : 24 May 2022 18:54
Operator : JC\MD
Sample : N2981-21
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TGMP-1

Quant Time: May 25 09:21:04 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N052422W.M
Quant Title : SW846 8260
QLast Update : Wed May 25 09:15:20 2022
Response via : Initial Calibration

