

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052422\
 Data File : VN072596.D
 Acq On : 24 May 2022 19:18
 Operator : JC\MD
 Sample : N2981-22
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TGMP-3

Quant Time: May 25 09:21:18 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N052422W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 25 09:15:20 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	132942	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	246704	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	346545	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.669	152	143234	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	83899	55.480	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 110.960%	
35) Dibromofluoromethane	8.016	113	73301	45.843	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 91.680%	
50) Toluene-d8	10.439	98	382043	62.575	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 125.140%	
62) 4-Bromofluorobenzene	12.727	95	149218	75.146	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 150.300%#	
Target Compounds						
16) Acetone	4.293	43	5866	7.892	ug/l	Qvalue 89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052422\
Data File : VN072596.D
Acq On : 24 May 2022 19:18
Operator : JC\MD
Sample : N2981-22
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TGMP-3

Quant Time: May 25 09:21:18 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N052422W.M
Quant Title : SW846 8260
QLast Update : Wed May 25 09:15:20 2022
Response via : Initial Calibration

