

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052422\
 Data File : VN072598.D
 Acq On : 24 May 2022 20:06
 Operator : JC\MD
 Sample : N2981-24
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TGMP-2

Quant Time: May 25 09:21:44 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N052422W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 25 09:15:20 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	158582	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	286793	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	337164	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.669	152	143207	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	91122	50.514	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	101.020%
35) Dibromofluoromethane	8.016	113	85889	46.207	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	92.420%
50) Toluene-d8	10.439	98	363744	51.250	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	102.500%
62) 4-Bromofluorobenzene	12.727	95	146709	63.555	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	127.120%
Target Compounds						
						Qvalue
16) Acetone	4.287	43	5996	6.762	ug/l #	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052422\
Data File : VN072598.D
Acq On : 24 May 2022 20:06
Operator : JC\MD
Sample : N2981-24
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TGMP-2

Quant Time: May 25 09:21:44 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N052422W.M
Quant Title : SW846 8260
QLast Update : Wed May 25 09:15:20 2022
Response via : Initial Calibration

