

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052422\
 Data File : VN072601.D
 Acq On : 24 May 2022 21:19
 Operator : JC\MD
 Sample : N2968-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-45

Quant Time: May 25 09:22:23 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N052422W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 25 09:15:20 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	159755	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	290736	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	326873	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.669	152	125770	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	92308	50.796	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 101.600%	
35) Dibromofluoromethane	8.016	113	85058	45.139	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 90.280%	
50) Toluene-d8	10.439	98	368395	51.201	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 102.400%	
62) 4-Bromofluorobenzene	12.727	95	139651	59.677	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 119.360%	
Target Compounds						
					Qvalue	
16) Acetone	4.293	43	2184	2.445	ug/l #	82
19) Methyl tert-butyl Ether	5.598	73	2282	0.458	ug/l #	54
25) 2-Butanone	7.334	43	1219	0.809	ug/l #	45

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052422\
Data File : VN072601.D
Acq On : 24 May 2022 21:19
Operator : JC\MD
Sample : N2968-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-45

Quant Time: May 25 09:22:23 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N052422W.M
Quant Title : SW846 8260
QLast Update : Wed May 25 09:15:20 2022
Response via : Initial Calibration

