

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052422\
 Data File : VN072607.D
 Acq On : 24 May 2022 23:44
 Operator : JC\MD
 Sample : N2968-08
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-44

Quant Time: May 25 09:24:10 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N052422W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 25 09:15:20 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	129126	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	234026	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	351697	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.669	152	116284	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	81971	55.807	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	111.620%
35) Dibromofluoromethane	8.016	113	70511	46.487	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	92.980%
50) Toluene-d8	10.434	98	356474	61.550	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	123.100%
62) 4-Bromofluorobenzene	12.727	95	122690	65.134	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	130.260%
Target Compounds						
						Qvalue
11) Tert butyl alcohol	5.357	59	15108	54.623	ug/l #	82
16) Acetone	4.287	43	2695	3.733	ug/l	100
19) Methyl tert-butyl Ether	5.604	73	22542	5.600	ug/l	100
22) Diisopropyl ether	6.487	45	5420	1.204	ug/l #	85
40) Benzene	8.457	78	1759	0.254	ug/l	98
95) Naphthalene	15.504	128	9278	1.282	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052422\
Data File : VN072607.D
Acq On : 24 May 2022 23:44
Operator : JC\MD
Sample : N2968-08
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-44

Quant Time: May 25 09:24:10 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N052422W.M
Quant Title : SW846 8260
QLast Update : Wed May 25 09:15:20 2022
Response via : Initial Calibration

