

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052422\
 Data File : VN072612.D
 Acq On : 25 May 2022 01:45
 Operator : JC\MD
 Sample : N3026-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 N48967

Quant Time: May 25 09:25:41 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N052422W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 25 09:15:20 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.080	168	125107	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	222131	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	241087	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	104137	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	83911	58.963	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	117.920%
35) Dibromofluoromethane	8.016	113	67699	47.023	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	94.040%
50) Toluene-d8	10.439	98	458294	83.368	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	166.740%#
62) 4-Bromofluorobenzene	12.727	95	102922	57.565	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	115.140%
Target Compounds						
					Qvalue	
16) Acetone	4.281	43	2911	4.162	ug/l #	88
40) Benzene	8.457	78	5974	0.909	ug/l #	89
52) Toluene	10.504	92	1408	0.338	ug/l	90
67) Ethyl Benzene	11.839	91	9941	1.083	ug/l	97
68) m/p-Xylenes	11.951	106	5373	1.535	ug/l	98
69) o-Xylene	12.274	106	1528	0.452	ug/l	93
78) n-propylbenzene	12.915	91	3165	0.308	ug/l	92
84) 1,2,4-Trimethylbenzene	13.362	105	9037	1.277	ug/l	91
95) Naphthalene	15.503	128	23825	3.676	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052422\
Data File : VN072612.D
Acq On : 25 May 2022 01:45
Operator : JC\MD
Sample : N3026-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
N48967

Quant Time: May 25 09:25:41 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N052422W.M
Quant Title : SW846 8260
QLast Update : Wed May 25 09:15:20 2022
Response via : Initial Calibration

