

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052423\
 Data File : VN077856.D
 Acq On : 24 May 2023 16:43
 Operator : JC\MD
 Sample : 02880-20
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 EB01-20230519

Quant Time: May 25 01:43:14 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 04:07:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	396318	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	713178	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	622934	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	199058	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	267525	45.114	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	90.220%		
35) Dibromofluoromethane	8.177	113	229328	49.312	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	98.620%		
50) Toluene-d8	10.576	98	868342	48.510	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	97.020%		
62) 4-Bromofluorobenzene	12.859	95	285053	45.089	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	90.180%		
Target Compounds						
30) Chloroform	7.982	83	12665	1.309	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052423\
Data File : VN077856.D
Acq On : 24 May 2023 16:43
Operator : JC\MD
Sample : 02880-20
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
EB01-20230519

Quant Time: May 25 01:43:14 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
Quant Title : SW846 8260
QLast Update : Tue May 16 04:07:42 2023
Response via : Initial Calibration

