

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052423\
 Data File : VN077859.D
 Acq On : 24 May 2023 17:55
 Operator : JC\MD
 Sample : VIBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VIBLK

Quant Time: May 25 01:43:45 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 04:07:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Pentafluorobenzene	8.236	168	314184	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.112	114	567633	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.877	117	492371	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	162868	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	213784	45.476	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	90.960%
35) Dibromofluoromethane	8.177	113	183134	49.476	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.960%
50) Toluene-d8	10.577	98	704728	49.464	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.920%
62) 4-Bromofluorobenzene	12.853	95	228693	45.449	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.900%
Target Compounds						
27) cis-1,2-Dichloroethene	7.494	96	12232	2.748	ug/l	97
44) Trichloroethene	9.359	130	70333	17.386	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052423\
 Data File : VN077859.D
 Acq On : 24 May 2023 17:55
 Operator : JC\MD
 Sample : VIBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VIBLK

Quant Time: May 25 01:43:45 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 04:07:42 2023
 Response via : Initial Calibration

