

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052423\
 Data File : VN077871.D
 Acq On : 25 May 2023 01:00
 Operator : JC\MD
 Sample : 02884-01
 Misc : 5.01g/10mL/100uL/5.0mL/MSVOA_N/MEOH
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 2001

Quant Time: May 25 05:00:43 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 04:07:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	353337	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	630812	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	553852	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	227719	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	232561	43.989	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	87.980%
35) Dibromofluoromethane	8.171	113	196775	47.837	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.680%
50) Toluene-d8	10.571	98	766089	48.386	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	96.780%
62) 4-Bromofluorobenzene	12.859	95	255103	45.620	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.240%
Target Compounds						
					Qvalue	
16) Acetone	4.453	43	9527	5.728	ug/l	93
80) 1,3,5-Trimethylbenzene	13.176	105	23863	1.446	ug/l	97
84) 1,2,4-Trimethylbenzene	13.488	105	101886	6.279	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052423\
Data File : VN077871.D
Acq On : 25 May 2023 01:00
Operator : JC\MD
Sample : 02884-01
Misc : 5.01g/10mL/100uL/5.0mL/MSVOA_N/MEOH
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
2001

Quant Time: May 25 05:00:43 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
Quant Title : SW846 8260
QLast Update : Tue May 16 04:07:42 2023
Response via : Initial Calibration

