

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052423\
 Data File : VN077874.D
 Acq On : 25 May 2023 02:11
 Operator : JC\MD
 Sample : 02852-02ME 20X
 Misc : 6.60g/5.0mL/100uL/5.0mL/MSVOA_N/MEOH
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 CB-32-G3-SO-14.5-051823ME

Quant Time: May 25 05:01:34 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 04:07:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	551612	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	969996	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	852963	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	328537	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	333393	40.394	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	80.780%
35) Dibromofluoromethane	8.177	113	286132	45.237	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.480%
50) Toluene-d8	10.577	98	1163528	47.791	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	95.580%
62) 4-Bromofluorobenzene	12.853	95	375600	43.681	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	87.360%
Target Compounds						
					Qvalue	
27) cis-1,2-Dichloroethene	7.494	96	22577	2.889	ug/l	94
44) Trichloroethene	9.359	130	136273	19.713	ug/l	99
86) p-Isopropyltoluene	13.735	119	223531	9.495	ug/l	99
95) Naphthalene	15.653	128	176356	7.852	ug/l #	74

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052423\
Data File : VN077874.D
Acq On : 25 May 2023 02:11
Operator : JC\MD
Sample : 02852-02ME 20X
Misc : 6.60g/5.0mL/100uL/5.0mL/MSVOA_N/MEOH
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
CB-32-G3-SO-14.5-051823ME

Quant Time: May 25 05:01:34 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
Quant Title : SW846 8260
QLast Update : Tue May 16 04:07:42 2023
Response via : Initial Calibration

