

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052423\
 Data File : VN077875.D
 Acq On : 25 May 2023 02:35
 Operator : JC\MD
 Sample : 02852-03ME 10X
 Misc : 6.95g/5.0mL/100uL/5.0mL/MSVOA_N/MEOH
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 CB-32-G3-SO-18-051823ME

Quant Time: May 25 05:01:48 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 04:07:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	477801	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	848873	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	745603	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	277388	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	299422	41.882	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	83.760%
35) Dibromofluoromethane	8.171	113	260613	47.081	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.160%
50) Toluene-d8	10.571	98	1032687	48.469	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	96.940%
62) 4-Bromofluorobenzene	12.853	95	328451	43.648	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	87.300%
Target Compounds						
					Qvalue	
27) cis-1,2-Dichloroethene	7.488	96	5358	0.792	ug/l	86
44) Trichloroethene	9.359	130	43123	7.128	ug/l	98
86) p-Isopropyltoluene	13.735	119	53206	2.677	ug/l	97
95) Naphthalene	15.653	128	37103	1.957	ug/l #	75

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052423\
Data File : VN077875.D
Acq On : 25 May 2023 02:35
Operator : JC\MD
Sample : 02852-03ME 10X
Misc : 6.95g/5.0mL/100uL/5.0mL/MSVOA_N/MEOH
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
CB-32-G3-SO-18-051823ME

Quant Time: May 25 05:01:48 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
Quant Title : SW846 8260
QLast Update : Tue May 16 04:07:42 2023
Response via : Initial Calibration

