

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052423\
 Data File : VN077881.D
 Acq On : 25 May 2023 04:59
 Operator : JC\MD
 Sample : 02852-09 10X
 Misc : 6.45g/5.0mL/100uL/5.0mL/MSVOA_N/MEOH
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 CB-32-G3-SO-50.5-051823

Quant Time: May 25 05:50:01 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 04:07:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	434593	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	804987	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	702335	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	230587	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	283992	43.673	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	87.340%
35) Dibromofluoromethane	8.171	113	245049	46.683	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.360%
50) Toluene-d8	10.577	98	974370	48.225	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	96.440%
62) 4-Bromofluorobenzene	12.853	95	299799	42.013	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	84.020%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052423\
Data File : VN077881.D
Acq On : 25 May 2023 04:59
Operator : JC\MD
Sample : 02852-09 10X
Misc : 6.45g/5.0mL/100uL/5.0mL/MSVOA_N/MEOH
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
CB-32-G3-SO-50.5-051823

Quant Time: May 25 05:50:01 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
Quant Title : SW846 8260
QLast Update : Tue May 16 04:07:42 2023
Response via : Initial Calibration

