

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052423\
 Data File : VN077883.D
 Acq On : 25 May 2023 05:47
 Operator : JC\MD
 Sample : 02852-11 10X
 Misc : 6.81g/5.0mL/100uL/5.0mL/MSVOA_N/MEOH
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 CB-32-G3-SO-55.5-051823

Quant Time: May 25 08:33:46 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 04:07:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	427199	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	780699	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	698400	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	240058	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	283852	44.407	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	88.820%
35) Dibromofluoromethane	8.171	113	243412	47.814	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.620%
50) Toluene-d8	10.571	98	958657	48.923	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.840%
62) 4-Bromofluorobenzene	12.859	95	301733	43.599	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	87.200%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052423\
Data File : VN077883.D
Acq On : 25 May 2023 05:47
Operator : JC\MD
Sample : 02852-11 10X
Misc : 6.81g/5.0mL/100uL/5.0mL/MSVOA_N/MEOH
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
CB-32-G3-SO-55.5-051823

Quant Time: May 25 08:33:46 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
Quant Title : SW846 8260
QLast Update : Tue May 16 04:07:42 2023
Response via : Initial Calibration

