

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052523\
 Data File : VN077899.D
 Acq On : 25 May 2023 14:03
 Operator : JC\MD
 Sample : 02850-05
 Misc : 1.05g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TRE-1729

Quant Time: May 25 16:30:14 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 04:07:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	410990	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	734796	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	662213	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	267749	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	265591	43.189	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	86.380%
35) Dibromofluoromethane	8.171	113	228192	47.624	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.240%
50) Toluene-d8	10.571	98	909445	49.311	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.620%
62) 4-Bromofluorobenzene	12.859	95	337535	51.820	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	103.640%
Target Compounds						
					Qvalue	
16) Acetone	4.453	43	41767	21.588	ug/l #	88
52) Toluene	10.635	92	261407	20.303	ug/l	97
84) 1,2,4-Trimethylbenzene	13.488	105	53026	2.780	ug/l	95
89) n-Butylbenzene	14.064	91	18200	1.031	ug/l #	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052523\
Data File : VN077899.D
Acq On : 25 May 2023 14:03
Operator : JC\MD
Sample : 02850-05
Misc : 1.05g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TRE-1729

Quant Time: May 25 16:30:14 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
Quant Title : SW846 8260
QLast Update : Tue May 16 04:07:42 2023
Response via : Initial Calibration

