

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052523\
 Data File : VN077903.D
 Acq On : 25 May 2023 15:38
 Operator : JC\MD
 Sample : 02944-13
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-VPB-RW11-EB-20230524

Quant Time: May 26 04:33:18 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 04:07:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	417000	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	773345	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	667155	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	213859	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	280740	44.995	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	89.980%
35) Dibromofluoromethane	8.177	113	239671	47.526	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.060%
50) Toluene-d8	10.571	98	936366	48.240	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	96.480%
62) 4-Bromofluorobenzene	12.859	95	308618	45.018	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.040%
Target Compounds						
30) Chloroform	7.971	83	12525	1.231	ug/l	Qvalue # 79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052523\
Data File : VN077903.D
Acq On : 25 May 2023 15:38
Operator : JC\MD
Sample : 02944-13
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-RW11-EB-20230524

Quant Time: May 26 04:33:18 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
Quant Title : SW846 8260
QLast Update : Tue May 16 04:07:42 2023
Response via : Initial Calibration

