

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052523\
 Data File : VN077905.D
 Acq On : 25 May 2023 16:26
 Operator : JC\MD
 Sample : 02944-03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-VPB-RW11-GW-338-340

Quant Time: May 26 04:34:14 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 04:07:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	422369	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.112	114	766307	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.877	117	656394	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	207258	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	284373	44.998	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	90.000%
35) Dibromofluoromethane	8.177	113	244860	49.001	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.000%
50) Toluene-d8	10.577	98	925307	48.108	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	96.220%
62) 4-Bromofluorobenzene	12.853	95	302266	44.497	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.000%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052523\
Data File : VN077905.D
Acq On : 25 May 2023 16:26
Operator : JC\MD
Sample : 02944-03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-RW11-GW-338-340

Quant Time: May 26 04:34:14 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
Quant Title : SW846 8260
QLast Update : Tue May 16 04:07:42 2023
Response via : Initial Calibration

