

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052523\
 Data File : VN077909.D
 Acq On : 25 May 2023 18:02
 Operator : JC\MD
 Sample : 02944-12
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-VPB-RW11-GW-458-460

Quant Time: May 26 04:36:06 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 04:07:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.236	168	372000	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	683412	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	592738	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	184837	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	259786	46.673	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	93.340%
35) Dibromofluoromethane	8.177	113	217273	48.755	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.500%
50) Toluene-d8	10.577	98	837396	48.819	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.640%
62) 4-Bromofluorobenzene	12.859	95	265555	43.834	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	87.660%
Target Compounds						
16) Acetone	4.459	43	4373	2.497	ug/l	Qvalue # 78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052523\
Data File : VN077909.D
Acq On : 25 May 2023 18:02
Operator : JC\MD
Sample : 02944-12
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-RW11-GW-458-460

Quant Time: May 26 04:36:06 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
Quant Title : SW846 8260
QLast Update : Tue May 16 04:07:42 2023
Response via : Initial Calibration

