

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052523\
 Data File : VN077912.D
 Acq On : 25 May 2023 19:14
 Operator : JC\MD
 Sample : 02898-04 100X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 DUP01-20230522

Quant Time: May 26 04:37:50 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 04:07:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.236	168	389167	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.112	114	720659	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	635383	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	204623	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.589	65	267728	45.978	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	91.960%
35) Dibromofluoromethane	8.171	113	222352	47.316	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.640%
50) Toluene-d8	10.577	98	880333	48.669	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.340%
62) 4-Bromofluorobenzene	12.853	95	287928	45.071	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.140%
Target Compounds						
					Qvalue	
4) Vinyl Chloride	2.524	62	31746	5.685	ug/l	96
27) cis-1,2-Dichloroethene	7.495	96	236946	42.979	ug/l	97
44) Trichloroethene	9.359	130	899996	175.234	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052523\
Data File : VN077912.D
Acq On : 25 May 2023 19:14
Operator : JC\MD
Sample : 02898-04 100X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
DUP01-20230522

Quant Time: May 26 04:37:50 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
Quant Title : SW846 8260
QLast Update : Tue May 16 04:07:42 2023
Response via : Initial Calibration

