

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052523\
 Data File : VN077916.D
 Acq On : 25 May 2023 20:49
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: May 26 04:40:41 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 04:07:42 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	90	0.00
2 T	Dichlorodifluoromethane	0.569	0.488	14.2	73	0.00
3 P	Chloromethane	0.682	0.588	13.8	86	0.00
4 C	Vinyl Chloride	0.717	0.732	-2.1#	90	0.01
5 T	Bromomethane	0.531	0.547	-3.0	98	0.00
6 T	Chloroethane	0.517	0.529	-2.3	96	0.00
7 T	Trichlorofluoromethane	1.000	1.019	-1.9	92	0.00
8 T	Diethyl Ether	0.365	0.389	-6.6	98	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.553	0.524	5.2	87	0.00
10 T	Methyl Iodide	0.660	0.820	-24.2	111	0.00
11 T	Tert butyl alcohol	0.103	0.096	6.8	86	0.00
12 CM	1,1-Dichloroethene	0.526	0.555	-5.5#	95	0.00
13 T	Acrolein	0.073	0.063	13.7	83	0.00
14 T	Allyl chloride	0.773	0.825	-6.7	104	0.00
15 T	Acrylonitrile	0.276	0.271	1.8	91	0.01
16 T	Acetone	0.235	0.206	12.3	85	0.00
17 T	Carbon Disulfide	1.417	1.364	3.7	92	0.01
18 T	Methyl Acetate	0.987	0.967	2.0	91	0.00
19 T	Methyl tert-butyl Ether	1.904	2.013	-5.7	95	0.00
20 T	Methylene Chloride	0.705	0.684	3.0	101	0.00
21 T	trans-1,2-Dichloroethene	0.591	0.635	-7.4	98	0.00
22 T	Diisopropyl ether	1.811	1.883	-4.0	95	0.00
23 T	Vinyl Acetate	1.122	1.126	-0.4	91	0.00
24 P	1,1-Dichloroethane	1.109	1.212	-9.3	99	0.00
25 T	2-Butanone	0.368	0.342	7.1	88	0.00
26 T	2,2-Dichloropropane	0.943	0.906	3.9	82	0.00
27 T	cis-1,2-Dichloroethene	0.708	0.778	-9.9	99	0.01
28 T	Bromochloromethane	0.480	0.485	-1.0	91	0.00
29 T	Tetrahydrofuran	0.241	0.226	6.2	88	0.00
30 C	Chloroform	1.220	1.288	-5.6#	99	0.00
31 T	Cyclohexane	1.017	0.900	11.5	85	0.00
32 T	1,1,1-Trichloroethane	1.038	1.138	-9.6	97	0.00
33 S	1,2-Dichloroethane-d4	0.748	0.740	1.1	89	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	87	0.00
35 S	Dibromofluoromethane	0.326	0.358	-9.8	93	0.00
36 T	1,1-Dichloropropene	0.477	0.515	-8.0	92	0.00
37 T	Ethyl Acetate	0.425	0.413	2.8	89	0.00
38 T	Carbon Tetrachloride	0.482	0.557	-15.6	98	0.00
39 T	Methylcyclohexane	0.584	0.536	8.2	78	0.00
40 TM	Benzene	1.407	1.600	-13.7	99	0.00
41 T	Methacrylonitrile	0.226	0.216	4.4	86	0.00
42 TM	1,2-Dichloroethane	0.511	0.550	-7.6	97	0.00
43 T	Isopropyl Acetate	0.694	0.708	-2.0	90	0.00
44 TM	Trichloroethene	0.356	0.411	-15.4	101	0.00
45 C	1,2-Dichloropropane	0.359	0.410	-14.2#	98	0.00
46 T	Dibromomethane	0.257	0.285	-10.9	100	0.00
47 T	Bromodichloromethane	0.492	0.568	-15.4	98	0.00
48 T	Methyl methacrylate	0.329	0.337	-2.4	89	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052523\
 Data File : VN077916.D
 Acq On : 25 May 2023 20:49
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: May 26 04:40:41 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 04:07:42 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.006	0.007	-16.7	97	0.00
50 S	Toluene-d8	1.255	1.309	-4.3	88	0.00
51 T	4-Methyl-2-Pentanone	0.413	0.424	-2.7	89	0.00
52 CM	Toluene	0.876	1.007	-15.0#	98	0.00
53 T	t-1,3-Dichloropropene	0.521	0.598	-14.8	96	0.00
54 T	cis-1,3-Dichloropropene	0.574	0.649	-13.1	95	0.00
55 T	1,1,2-Trichloroethane	0.348	0.402	-15.5	102	0.00
56 T	Ethyl methacrylate	0.526	0.585	-11.2	94	0.00
57 T	1,3-Dichloropropane	0.596	0.679	-13.9	98	0.00
58 T	2-Chloroethyl Vinyl ether	0.259	0.230	11.2	77	0.00
59 T	2-Hexanone	0.305	0.298	2.3	85	0.00
60 T	Dibromochloromethane	0.359	0.431	-20.1	100	0.00
61 T	1,2-Dibromoethane	0.347	0.408	-17.6	98	0.00
62 S	4-Bromofluorobenzene	0.443	0.446	-0.7	87	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	86	0.00
64 T	Tetrachloroethene	0.350	0.433	-23.7	110	0.00
65 PM	Chlorobenzene	1.084	1.230	-13.5	99	0.00
66 T	1,1,1,2-Tetrachloroethane	0.375	0.465	-24.0	103	0.00
67 C	Ethyl Benzene	1.942	2.154	-10.9#	94	0.00
68 T	m/p-Xylenes	0.736	0.843	-14.5	97	0.00
69 T	o-Xylene	0.724	0.832	-14.9	97	0.00
70 T	Styrene	1.150	1.328	-15.5	96	0.00
71 P	Bromoform	0.250	0.306	-22.4	97	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	83	0.00
73 T	Isopropylbenzene	4.310	4.936	-14.5	92	0.00
74 T	N-amyl acetate	1.593	1.619	-1.6	85	0.00
75 P	1,1,2,2-Tetrachloroethane	1.326	1.415	-6.7	91	0.00
76 T	1,2,3-Trichloropropane	1.252	1.234	1.4	83	0.00
77 T	Bromobenzene	0.984	1.124	-14.2	97	0.00
78 T	n-propylbenzene	5.081	5.509	-8.4	89	0.00
79 T	2-Chlorotoluene	3.032	3.438	-13.4	94	0.00
80 T	1,3,5-Trimethylbenzene	3.622	4.122	-13.8	94	0.00
81 T	trans-1,4-Dichloro-2-butene	0.392	0.402	-2.6	84	0.00
82 T	4-Chlorotoluene	3.017	3.309	-9.7	93	0.00
83 T	tert-Butylbenzene	3.162	3.431	-8.5	88	0.00
84 T	1,2,4-Trimethylbenzene	3.563	4.125	-15.8	95	0.00
85 T	sec-Butylbenzene	4.494	4.572	-1.7	82	0.00
86 T	p-Isopropyltoluene	3.583	3.862	-7.8	86	0.00
87 T	1,3-Dichlorobenzene	1.819	1.961	-7.8	92	0.00
88 T	1,4-Dichlorobenzene	1.805	1.929	-6.9	91	0.00
89 T	n-Butylbenzene	3.296	3.103	5.9	76	0.00
90 T	Hexachloroethane	0.589	0.641	-8.8	83	0.00
91 T	1,2-Dichlorobenzene	1.699	1.978	-16.4	96	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.257	0.245	4.7	81	0.00
93 T	1,2,4-Trichlorobenzene	0.993	0.839	15.5	71	0.00
94 T	Hexachlorobutadiene	0.412	0.345	16.3	67	0.00
95 T	Naphthalene	3.418	2.771	18.9	69	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052523\
Data File : VN077916.D
Acq On : 25 May 2023 20:49
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: May 26 04:40:41 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
Quant Title : SW846 8260
QLast Update : Tue May 16 04:07:42 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.972	0.862	11.3	73	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6