

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052523\
 Data File : VN077907.D
 Acq On : 25 May 2023 17:14
 Operator : JC\MD
 Sample : 02944-10
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-VPB-RW11-DUP-20230523

Quant Time: May 26 04:35:06 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 04:07:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	415946	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.112	114	770590	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	663782	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	213740	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	283342	45.527	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	91.060%
35) Dibromofluoromethane	8.177	113	243028	48.364	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.720%
50) Toluene-d8	10.576	98	939642	48.582	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.160%
62) 4-Bromofluorobenzene	12.859	95	304249	44.540	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.080%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052523\
Data File : VN077907.D
Acq On : 25 May 2023 17:14
Operator : JC\MD
Sample : 02944-10
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-RW11-DUP-20230523

Quant Time: May 26 04:35:06 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
Quant Title : SW846 8260
QLast Update : Tue May 16 04:07:42 2023
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.235 min Scan# 110
Delta R.T. 0.005 min
Lab File: VN077907.D
Acq: 25 May 2023 17:14




Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-RW11-DUP-20230523


Tgt Ion:168 Resp: 415946
Ion Ratio Lower Upper
168 100
99 62.9 53.0 79.6



#33
1,2-Dichloroethane-d4
Concen: 45.527 ug/l
RT: 8.588 min Scan# 1128
Delta R.T. 0.006 min
Lab File: VN077907.D
Acq: 25 May 2023 17:14

Tgt Ion: 65 Resp: 283342
Ion Ratio Lower Upper
65 100
67 52.2 0.0 106.6





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.112 min Scan# 11

Delta R.T. 0.006 min

Lab File: VN077907.D

Acq: 25 May 2023 17:14

Instrument :

MSVOA_N

ClientSampleId :

BP-VPB-RW11-DUP-20230523





Tgt Ion:114 Resp: 770590

Ion Ratio Lower Upper

114 100

63 19.6 0.0 36.4

88 15.2 0.0 30.6



#35

Dibromofluoromethane

Concen: 48.364 ug/l

RT: 8.177 min Scan# 1058

Delta R.T. 0.006 min

Lab File: VN077907.D

Acq: 25 May 2023 17:14




Tgt Ion:113 Resp: 243028


Ion Ratio Lower Upper

113 100

111 103.4 82.6 123.8

192 16.1 12.7 19.1









#50
Toluene-d8
Concen: 48.582 ug/l
RT: 10.576 min Scan# 1465
Delta R.T. 0.005 min
Lab File: VN077907.D
Acq: 25 May 2023 17:14




Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-RW11-DUP-20230523


Tgt Ion: 98 Resp: 939642
Ion Ratio Lower Upper
98 100
100 62.5 52.0 78.0



#62
4-Bromofluorobenzene
Concen: 44.540 ug/l
RT: 12.859 min Scan# 1854
Delta R.T. 0.006 min
Lab File: VN077907.D
Acq: 25 May 2023 17:14

Tgt Ion: 95 Resp: 304249
Ion Ratio Lower Upper
95 100
174 72.9 0.0 143.8
176 69.1 0.0 135.8





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.870 min Scan# 1

Delta R.T. 0.000 min

Lab File: VN077907.D

Acq: 25 May 2023 17:14

Instrument :

MSVOA_N

ClientSampleId :

BP-VPB-RW11-DUP-20230523





Tgt Ion:117 Resp: 663782

Ion Ratio Lower Upper

117 100

82 55.3 44.0 66.0

119 32.1 25.6 38.4



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.800 min Scan# 2014

Delta R.T. -0.000 min

Lab File: VN077907.D

Acq: 25 May 2023 17:14

Tgt Ion:152 Resp: 213740

Ion Ratio Lower Upper

152 100

115 63.6 32.3 96.8

150 158.5 0.0 355.8

