

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052621\
 Data File : VN067168.D
 Acq On : 26 May 2021 14:12
 Operator : JC/MD
 Sample : VN0526WBL01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0526WBL01

Quant Time: May 27 02:38:59 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N052121W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 21 17:24:08 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.597	168	227761	50.00	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.525	114	428021	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.357	117	368789	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.302	152	126378	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.959	65	175801	49.74	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	99.48%
35) Dibromofluoromethane	7.519	113	132691	50.52	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	101.04%
50) Toluene-d8	10.035	98	499196	49.38	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	98.76%
62) 4-Bromofluorobenzene	12.360	95	157728	45.64	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	91.28%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052621\
Data File : VN067168.D
Acq On : 26 May 2021 14:12
Operator : JC/MD
Sample : VN0526WBL01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0526WBL01

Quant Time: May 27 02:38:59 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N052121W.M
Quant Title : SW846 8260
QLast Update : Fri May 21 17:24:08 2021
Response via : Initial Calibration

