

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052622\
 Data File : VN072636.D
 Acq On : 26 May 2022 17:32
 Operator : JC\MD
 Sample : VN0526WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0526WBL01

Quant Time: May 27 05:33:59 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N052622W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri May 27 05:07:24 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.651	128	29957	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	8.963	114	136976	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.745	117	86768	30.000	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.434	65	82164	28.663	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	95.533%
60) 4-Bromofluorobenzene	12.727	95	48166	31.609	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	105.367%
63) Toluene-d8	10.439	98	162656	32.969	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	109.900%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052622\
Data File : VN072636.D
Acq On : 26 May 2022 17:32
Operator : JC\MD
Sample : VN0526WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0526WBL01

Quant Time: May 27 05:33:59 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N052622W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Fri May 27 05:07:24 2022
Response via : Initial Calibration

