

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052622\
 Data File : VN072640.D
 Acq On : 26 May 2022 20:17
 Operator : JC\MD
 Sample : N3038-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TT-045-IDWGW-20220524

Manual Integrations
 APPROVED

Reviewed By : John Carlone 05/27/2022
 Supervised By : Mahesh Dadoda 05/27/2022

Quant Time: May 27 05:35:16 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N052622W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri May 27 05:07:24 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.651	128	31922	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	8.963	114	171787	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.739	117	138402	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.434	65	90282	29.556	ug/l	0.00
Spiked Amount 30.000	Range 91 - 110		Recovery =	98.533%		
60) 4-Bromofluorobenzene	12.727	95	56621	23.295	ug/l	0.00
Spiked Amount 30.000	Range 63 - 112		Recovery =	77.667%		
63) Toluene-d8	10.439	98	248396	31.564	ug/l	0.00
Spiked Amount 30.000	Range 91 - 112		Recovery =	105.200%		
Target Compounds						Qvalue
15) Acetone	4.298	58	4378m	10.465	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052622\
Data File : VN072640.D
Acq On : 26 May 2022 20:17
Operator : JC\MD
Sample : N3038-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TT-045-IDWGW-20220524

Quant Time: May 27 05:35:16 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N052622W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Fri May 27 05:07:24 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :John Carlone 05/27/2022
Supervised By :Mahesh Dadoda 05/27/2022

